

Garantía Limitada

La información de garantía siguiente es un SUMARIO de la Garantía del Producto completa provisto de GF Health Products, Inc. Para la garantía completa, por favor de visitar a la página de web www.grahamfield.com

GF Health Products, Inc. (“GF”) garantiza al comprador original o arrendatario (cual quiera is el “Cliente”) solamente, que va reemplazar o reparar componentes, a la decisión solo de GF, cuales son defectos en material o calidad utilizado normalmente y con servicio. La garantía es para partes solamente y no incluye fuerza laboral o el costo de transporte. Todas garantías con condicionales por el uso apropiado del producto estrictamente en acuerdo con las instrucciones aplicables de GF. Producto próvido a usuarios como réntales deben ser serviciados por el proveedor y inspeccionado antes de entrega. Cada usuario debe ser preparado en la operación y seguridad de este producto antes de uso. Dentro las reglas normales en este documento, la estructura tiene garantía por cinco (5) anos y los componentes de uso normal como puntas, ruedas y manijas tienen garantía por tres (3) meses. El periodo de garantía se comienza en la fecha de transporte al Cliente a menos que haya una fecha de expiración en el componente en cual la garantía se expira en la periodo de garantía o la fecha de expiración.

OBTENIENDO SERVICIO DE GARANTIA

La garantía limitada solamente aplica a defectos reportados al Distribuidor de quien el Cliente compro el producto dentro el periodo de garantía aplicable. Si no hay Distribuidor, debes llamar a GF directamente llamando (678) 291-3207, mandando fax a (770)-368-2386 o por email a cs@grahamfield.com. Direcciones especificos se van a dar por el Representante de Servicio De Clientes.

EXCLUSIONES

La garantía no cubre y GF será responsable por lo siguiente: 1) Defectos, daño o otras condiciones causadas por mal uso, abuso, negligencia, alteración, accidente, daño de transporte o falta de buscar y obtener reparo o repuesto en manera puntual; 2) Productos no instalados, utilizados or mantenidos correctamente; y 3) Cargos de Labor o transporte. Otras exclusiones se aplican.

GARANTIA ENTERO, REMEDIO EXCLUSIVO Y DESCARGO DE DANOS CONSIGUIENTE

ESTA GARANTIA ES LA UNICA GARANTIA DE GF Y ES EN VEZ DE TODAS OTRAS GARANTIAS, EXPRESIDAS O IMPLICITADAS.

Para la garantía completa por este producto, por favor de visitar a la página www.grahamfield.com

LUMEX®

7910A-1/7915A-1

- Designed for transport to and from the shower
- Anodized, non-corrosive aluminum frame construction will not rust
- 5" swivel casters with rear wheel locks facilitate maneuverability in tight spaces
- Rolls easily over most toilets
- Allows complete access to body for easy cleaning during bathing
- Maximum Weight Capacity 300 lb - EVENLY DISTRIBUTED
- Lifetime Limited Warranty

7915A-1 Only

- Powder coated steel front brace with footrests
- Foot plates fold up and out of the way for unobstructed entry and exit

7910A-1/7915A-1

- Diseñado para transporte seguro a y de la ducha
- Marco construido de aluminio anodizado y no-corrosivo que es inoxidable
- Apoyabrazos almohada aumenta comodidad del usuario
- Ruedas giratorias de 5" con cerraduras en las ruedas traseras facilita maniobrabilidad en espacios apretados
- Permite acceso completo al cuerpo para limpiar fácilmente durante bañado
- Capacidad de peso 300 lb UNIFORMEMENTE DISTRIBUIDA
- Garantía Limitada de por Vida

7915A-1 solamente

- Abrazadera delantera de acero cubierto con reposapiés
- Reposapiés plataformas se doblan fuera del area para entrar y salir facilmente

For the most updated and current information on packaging, warranties, products and specifications, please visit our website at www.grahamfield.com

Para la información más ultimo sobre embalaje, garantias, productos y especificaciones, por favor de visitar a nuestro sitio de internet a www.grahamfield.com

© 2014, GF Health Products, Inc.
2935 Northeast Parkway, Atlanta, Georgia 30360
Made in China www.grahamfield.com

GF0500039RevD14

LUMEX®

Shower Transport Chair

Silla de Transporte de Ducha

Items 7910A-1 & 7915A-1



User Instructions and Warranty Information

Important: Read Before Use

Instrucciones e Información de la Garantía
Importante: Lea Antes de Uso



GRAHAM-FIELD

Limited Warranty

The following warranty information is a SUMMARY of the complete Product Warranty provided by GF Health Products, Inc. For the complete warranty, please visit our website at www.grahamfield.com.

GF Health Products, Inc. (“GF”) warrants to the original retail purchaser or lessee (either is the “Customer”) only, that it will replace or repair components, at GF’s sole discretion, that are defective in material or workmanship under normal use and service. The warranty is for parts only and does not include labor or costs of shipping. All warranties are conditioned upon the proper use of the product strictly in accordance with applicable GF instructions. Product provided to users as rentals must be serviced by the provider and inspected prior to delivery. Each user must be properly trained on the operation and safety of this product prior to use. Within the guidelines set forth in this document, the frame is warranted for five (5) years and the normal wear components such as tips, wheels/casters, and handgrips are warranted for three (3) months. The applicable warranty period shall commence from date of shipment to the Customer, unless there is an expiration date on the component in which case the warranty shall expire on the earlier of the warranty period or the expiration date.

OBTAINING WARRANTY SERVICE

This limited warranty shall only apply to defects that are reported to the Dealer from whom the Customer purchased the product within the applicable warranty period. If there is not a Dealer, you must contact GF directly by calling (678) 291-3207, sending a fax request to 770-368-2386 or by e-mailing a request to cs@grahamfield.com. Specific directions will be provided by the Customer Service Representative.

EXCLUSIONS

The warranty does not cover and GF shall not be liable for the following: 1) Defects, damage, or other conditions caused by misuse, abuse, negligence, alteration, accident, freight damage, tampering or failure to seek and obtain repair or replacement in a timely manner; 2) Products not installed, used, or properly cleaned and maintained; and 3) Labor or shipping charges. Other exclusions apply.

ENTIRE WARRANTY, EXCLUSIVE REMEDY AND CONSEQUENTIAL DAMAGES DISCLAIMER

THIS WARRANTY IS GF’S ONLY WARRANTY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED.

For the complete warranty for this product, please visit our website at www.grahamfield.com.

Instructions for Use – Please Read All Instructions Before Use.

Attaching the Backrest (7910A-1 & 7915A-1)

1. Remove chair and back support from the box. Place chair on a flat surface.
2. Lock rear casters to prevent chair from moving during assembly.
3. Locate the channel guides (2 each) on the back side of the chair frame near the seat. Align the backrest support tubes with the channel guides.
4. Depress the push buttons on the back support tubing and slide the tubing through the channel guides until the push buttons protrude fully through the channel guide holes.

Attaching the Footrests (Only for 7915A-1)

1. Locate the footrests (2 each).
 2. Insert 1 footrest into the chair leg extension.
 3. Ensure footplate is able to fold down to the inside.
 4. Move footrest upward or downward to desired length, ensuring alignment holes match with holes in chair leg extension.
 5. Once the holes are aligned, insert a screw into the hole, rotating the screw clockwise. Repeat this step for the second screw.
 6. Use a screw driver to tighten the screws.
- Caution: Do not over tighten.**
7. Repeat these steps for the remaining footrest.

Maintenance

- Use a mild soap and water solution, rinse with clear water, and then dry with a soft cloth. Ensure that chair is well rinsed before using.
- Never use organic solvents, abrasive cleaners, or scouring pads on any part of chair.
- Check fit and tightness of all nuts, bolts, and knobs periodically to ensure that chair is safe to use.
- To guarantee the best service from your chair, always specify Lumex replacement parts, available from your GF Health Products, Inc. distributor.

For more information regarding GF Health Products, Inc. products, parts and services, please contact your distributor or visit us online at www.grahamfield.com.

Safety Guidelines – Please Read Before Use

NOTE: Check ALL parts for shipping damage. If shipping damage is noted, DO NOT use. Contact carrier/dealer for further instruction.

WARNING ⚠

WARNING: GF Health Products Inc. assumes no responsibility for any damage or injury caused by improper installation, assembly or use of this product.

WARNING: This product has a maximum weight limit of 250lb. EVENLY DISTRIBUTED

WARNING: Users with limited physical capabilities should be supervised or assisted when using this product.

WARNING: Casters **MUST** be in contact with the floor at all times during use.

WARNING: Make sure that all attaching hardware, screws, nuts and/or bolts are tight at all times.

WARNING: This product should not be used without proper instruction from a healthcare professional.

WARNING: Notice for California Customers
California Proposition 65 WARNING: This product contains a chemical known to the State of California to cause cancer and reproductive or developmental harm.

WARNING: If components are damaged or missing, contact your GF Health Products, Inc. distributor immediately.



Instrucciones para Uso – Lea Todas Las Instrucciones Antes de Uso.

Sujetando el respaldo (para 7910A-1 y 7915A-1)

1. Remueva la silla y el soporte de espalda de la caja. Ponga la silla en una superficie plana.
2. Asegure las ruedas traseras para evitar silla se mueva durante el montaje.
3. Busque las guías de canal (2) en la parte posterior del marco de la silla cerca del asiento. Alinee los tubos de soporte de la espalda con las guías de canal.
4. Presione los botones en el tubo soporte de la espalda y deslice el tubo a través de las guías de canal hasta que los botones sobresalten completamente por los huecos de las guías de canal.

Sujetando los Reposapiés (Solamente para 7915A-1)

1. Localice los reposapiés (2).
2. Inserta un reposapié en la extensión de pata de la silla.
3. Asegúrese de que el reposapié se puede doblar hace adentro.
4. Mueve el reposapié hacia arriba y abajo hasta la altura deseada, asegurando de que los huecos de ajustamiento machean con los huecos en la extensión de pata de la silla.
5. Cuando están alineados los huecos, inserta un tornillo en el hueco, dando vuelta al tornillo hacia la derecha. Repita este paso para el tornillo segundo.
6. Utilice un destornillador para apretar los tornillos.
Precaución: No sobre apriete.
7. Repita estos pasos para el otro reposapié.



Mantenimiento

- Utilice una solución de jabón suave y agua, enjuague con agua limpia y seca con una toalla suave. Asegure que la silla de ducha esta enjuagado bien antes de uso.
- Nunca utilice solventes orgánicos, limpiadores abrasivos en cual quiere parte de este inodoro.
- Cheque seminalmente que los tornillos y perillas están apretados para asegurar que la silla de ducha esta bueno para utilizar.
- Para garantizar el servicio mejor de su inodoro, siempre pide partes de reemplace de Lumex si son disponible por so distribuidor de GF Health Products, Inc.

Para mas información sobre su producto, partes o servicio de GF Health Products, Inc, por favor de llamar a su distribuidor o visitenos online en www.grahamfield.com.

Directriz de Seguridad – Por Favor de Leer Antes de Uso

NOTA: Examine to das las partes por daño en transporte. Si hay daño en transporte, NO UTILICE. Llame a su empresa de transporte para instrucciones.

ADVERTENCIA ⚠

ADVERTENCIA: GF Health Products Inc. no asume ninguna responsabilidad por cual quiere daño o herida causado por instalación, asamblea o uso incorrecto de este producto.

ADVERTENCIA: Este producto tiene una capacidad de peso máximo de 250 lb. DISTRIBUIDO UNIFORMEMENTE

ADVERTENCIA: Usuarios con capacidades físico limitado deben ser supervisados o asistidos cuando usando la silla bañera.

ADVERTENCIA: Asegure que TODAS los tornillos y perillas están apretadas todo el tiempo.

ADVERTENCIA: Las ruedas deben estar en contacto con el suelo en todo momento durante su uso.

ADVERTENCIA: Este producto no debe ser usado sin instrucción apropiado de su profesional de cuidado de salud.

ADVERTENCIA: Aviso a Clientes de California
ADVERTENCIA Proposición 65 de California: Este producto contiene una cémica que el Estado de California de causar cancer y herida reproductiva y de desarrollo.

ADVERTENCIA: Si los componentes están dañados o si faltan, llame a su distribuidor de GF Health Products, Inc. inmediatamente.